



Anterior Chamber Maintenance Cannulas

For maintaining and forming the anterior chamber during surgery by injecting or removing viscoelastic materials, BSS or air.


Anterior Chamber Maintainers

Code	Gauge and size	Description	Tip
			
M2658	23G x 9 over 64"	0.64 x 3.5mm	Anterior Chamber Maintainer. 150mm silicone tubing. Micro-etched allowing cannula to be self retaining (Lewicky)
M2657	20G x 9 over 64"	0.9 x 3.5mm	Anterior Chamber Maintainer. 150mm silicone tubing. Micro-etched allowing cannula to be self retaining (Lewicky)

Anterior Chamber Maintenance Cannulas

Code	Gauge and size	Description	Tip
			
M2639	● 25G x 1"	0.5 x 25mm	Anterior Chamber Cannula. Angled 35°, 7.5mm from tip (Rycroft)
M2270	● 30G x 7/8"	0.3 x 22mm	Anterior Chamber Cannula. Angled 35°, 4mm from tip (Rycroft)
M2270A	● 30G x 7/8"	0.3 x 22mm	Anterior Chamber Cannula. Angled 35°, 8mm from tip (Rycroft)
M2273	● 30G x 7/8"	0.3 x 22mm	Anterior Chamber Cannula. Angled 35°, 5.5mm from tip (Rycroft)
M2273A	● 27G x 7/8"	0.4 x 22mm	Anterior Chamber Cannula. Angled 35°, 4.5mm from tip (Rycroft)
M2279	● 27G x 7/8"	0.4 x 22mm	Anterior Chamber Cannula. Angled 45°, 11mm from tip (Rycroft)
M2275	● 25G x 7/8"	0.5 x 22mm	Anterior Chamber Cannula. Angled 35°, 5mm from tip (Rycroft)
M2273G	● 23G x 7/8"	0.64 x 22mm	Anterior Chamber Cannula. Angled 35°, 5.5mm from tip (Rycroft)
M2642	● 21G x 7/8"	0.81 x 22mm	Anterior Chamber Cannula. Angled 35°, 7.5mm from tip (Rycroft)
M2273B	● 20G x 7/8"	0.9 x 22mm	Anterior Chamber Cannula. Angled 45°, 7.5mm from tip (Rycroft)

Wash Out Cannulas

Code	Gauge and size	Description	Tip
			
M9280	● 19G x 1"	1.1 x 25mm	Washout Cannula. Angled 32°, 8mm from tip (Bishop Harmon)